



**Trilogy 200™: Mode SIMV (Active PAP Circuit)**

Date (M/D/Y): \_\_\_\_\_

Client Name: \_\_\_\_\_

Client D.O.B (M/D/Y): \_\_\_\_\_

Client Address: \_\_\_\_\_

Serial Number: \_\_\_\_\_

Clinician Name:(print): \_\_\_\_\_ Clinician Signature: \_\_\_\_\_

Dual Prescription: On/Off  Primary Setting  Secondary Setting

Circuit Type: **Active PAP**

- Tidal Volume (ml) \_\_\_\_\_
- Breath Rate (BPM) \_\_\_\_\_
- Inspiratory Time (s) \_\_\_\_\_
- Flow Pattern \_\_\_\_\_
- PEEP \_\_\_\_\_
- Pressure Support Above PEEP \_\_\_\_\_
- Flow Trigger Sensitivity \_\_\_\_\_
- Flow Cycle Sensitivity \_\_\_\_\_
- Rise Time \_\_\_\_\_
- Sigh \_\_\_\_\_
- Nebulizer Enabled \_\_\_\_\_
- Circuit Disconnect (MPV) Alarm (s) \_\_\_\_\_
- Low Insp. Pressure Alarm (cmH20) \_\_\_\_\_
- High Insp. Pressure Alarm (cmH20) \_\_\_\_\_
- Apnea Alarm (s) \_\_\_\_\_
- Low Vti Alarm \_\_\_\_\_
- High Vti Alarm \_\_\_\_\_
- Low MV Alarm \_\_\_\_\_
- High MV Alarm \_\_\_\_\_
- Low RR Alarm \_\_\_\_\_
- High RR Alarm \_\_\_\_\_